


© 2004 Blackwell Publishing Ltd *Journal of Internal Medicine* 255: 103–110





An inspection certificate is supplied as standard.
Refer to page U-11 for details.



Dial Caliper SERIES 505

- Newly designed dial movement for ultra-smooth sliding.
- Easy-to-read yellow dial.
- Large finger-rest aids ease-of-use.
- Jaw tips are relieved for easy measurement of thin parts.
- Allows step measurement.



505-730




505-731



505-745

DIMENSIONS



Range (mm)	A	B	D	E	H	L
0 - 150	16.5	21	40	57.2	16	231
0 - 200	20	24.5	50			288
0 - 300	22	27.5	64	70.2	20	404


SPECIFICATIONS

Metric				
Order No.	Range (mm)	Graduation (mm)	Accuracy (mm)	Remarks
505-730	0 - 150	0.02, 2/rev	±0.03	Carbide-tipped jaws for outside measurement
505-734				Carbide-tipped jaws for outside and inside measurement
505-735		0.01, 1/rev	±0.02	
505-732*	0 - 200	0.02, 2/rev	±0.03	—
505-731		0.01, 1/rev		
505-733*		0.02, 2/rev	±0.04	
505-745	0 - 300	0.02, 2/rev	±0.04	

* Silver cover type

Inch				
Order No.	Range (in)	Graduation (in)	Accuracy (in)	Remarks
505-740J	0 - 6	0.001, 0.2/rev	±0.001	
505-742J*			±0.001	
505-742-56J			±0.001	
505-742-51J		0.001, 0.1/rev	±0.001	Carbide-tipped jaws for outside measurement
505-736*			±0.001	Carbide-tipped jaws for outside and inside measurement
505-738*	0 - 8	0.001, 0.2/rev	±0.001	Carbide-tipped jaws for outside measurement
505-744			±0.002	
505-741J		0.001, 0.1/rev	±0.002	—
505-743J*			±0.002	Carbide-tipped jaws for outside measurement
505-737*		0.001, 0.2/rev	±0.002	Carbide-tipped jaws for outside and inside measurement
505-739*	0 - 12	0.001, 0.2/rev	±0.002	
505-749			±0.002	
505-746*			±0.002	
505-750		0.001, 0.1/rev	±0.002	Carbide-tipped jaws for outside measurement
505-747*			±0.002	Carbide-tipped jaws for outside and inside measurement
505-748*		0.001, 0.1/rev	±0.002	Carbide-tipped jaws for outside and inside measurement

* Silver cover type



FOR MORE INFORMATION PLEASE CONTACT:



BHAGWATI

Hardware & Mill Store Since 1976

Address: 4769, 1ST Floor, Old Post Office Building,
Hauz Qazi, Delhi - 110006

Mobile: 9212012856, 9350076343

Website: www.bhagwathardware.com
www.bhagwathardwaredelhi.in

E-mail: sandeep@bhagwatimail.com

AN ISO 9001 : 2015 CERTIFIED COMPANY